

# pHG-2091Pro

## Online pH / ORP Meter

## User manual



**Shanghai BOQU Instrument Co., Ltd.**

Address: No. 118 Xiuyan Road, Pudong New Area, Shanghai, zip code: 201315, China

Web: [www.boquinstrument.com](http://www.boquinstrument.com) | Email: [michael@shboqu.com](mailto:michael@shboqu.com) | Mob: 86-15000087545

# Content

<b>Product Specification Sheet</b>	.....	1
<b>Installation and wiring</b>	.....	2
Size	.....	2
Install	.....	2
Wiring	.....	3
<b>Operation panel</b>	.....	4
<b>Measurement interface</b>	.....	5
<b>Set up</b>	.....	6
Set interface	.....	6
Measurement setup	.....	6
4-20mA	.....	7
Communication settings	.....	7
Temperature setting	.....	7
Relay 1	.....	8
Relay 2	.....	8
language settings	.....	9
Backlight setting	.....	9
reset	.....	9
<b>calibration</b>	.....	10
Automatic calibration	.....	10
Manual calibration	.....	11
Three-point calibration	.....	11
Buffer group	.....	11
Electrode state	.....	12
reset	.....	12
<b>appendix</b>	.....	13

# Product description

The PHG-2091Pro online pH ORP meter is a brand new online intelligent analog meter independently developed and manufactured by BOQU Instrument. It has a built-in A/D conversion module and is compatible with a variety of analog signal electrodes. Complete functions, stable performance, easy operation, low power consumption, safety and reliability are the outstanding advantages of this instrument. This instrument is equipped with RS485 transmission interface, which can be connected to the upper computer through Modbus RTU protocol to realize monitoring and recording.


This online pH and ORP meter uses a matching analog signal pH electrode and ORP electrode (see appendix 2 for details), which can be widely used in solutions, wastewater pH or ORP values and temperatures in industrial occasions such as thermal power generation, chemical industry, metallurgy, environmental protection, pharmaceuticals, biochemical, food, and tap water. Continuous monitoring.

# Product Specification Sheet


Specification	details
Name	Online pH/ORP meter
Measuring range	-2 ~16 pH -2000 ~2000 mV -30 ~130 °C
Measurement accuracy	±1%FS ±0.5°C
Shell material	ABS plastic
Power Supply	90 – 260V AC 50/60Hz
Power consumption	4W
Output method	Two 4-20mA outputs, RS485
Relay	5A/250V AC      5A/30V DC
Communication	RS485 Modbus RTU
Waterproof level	IP65
Size	98.2×98.2×128.3mm
Hole size	92*92mm

# Installation and wiring


## Instrument size chart




## Embedded installation diagram



## Pipeline installation diagram



## Wiring diagram



## 1. Operation panel

The main panel of the pH ORP controller has 2 modules, the LED LCD display module and the button module.

The user can set and adjust the parameters of the instrument through the 5 keys on the panel.




Figure 1 Measuring instrument operation panel

- ① Set/Exit button
- ② Select/change button
- ③ Up selection button
- ④ Down selection button
- ⑤ Confirm button
- ⑥ Instrument display

## 1. Measurement interface

After the start-up animation ends, enter the main measurement interface.

When the pH ORP meter is working normally, the LED display shows the following contents.




Figure 2 The main interface of measurement display

- ① Measurement reading
- ② Measurement unit
- ③ Measure temperature
- ④ pH/ORP The corresponding value of 4-20mA
- ⑤ Measurement mode
- ⑥ Over/low limit alarm prompt

Three, settings

Press the "Set/Exit button" to enter the password input interface.

PASSWORD
0000

Figure 3 Enter the password

**Enter settings:**

Enter the password "3700" to enter the setup menu.

SETUP	SETUP
1. Measure	3. RS485
2. 4-20mA	4. Temp
SETUP	SETUP
5. Relay1 Func.	7. Relay2
6. Relay1	8. Language
Factory Reset	
► Yes	
No	

Figure 4 Setting menu

### 3.1 Measurement settings

Under this menu, the user can change the measurement method pH / ORP.

Measure
► pH
ORP

Figure 3.1 Measurement settings

### 3.2 4-20mA setting

In this menu, the user can change the corresponding value of 4-20mA and set the corresponding effective range.

4-20mA	
Measure Value	Temp
4mA : 0000 pH 20mA : 1000 pH	4mA : +000 °C 20mA : +100 °C

Figure 3.2 4-20mA setting

### 3.3 Communication settings

In this menu, the user can change the communication address and communication speed.

RS485	
Address	► 4800 bps 9600 bps 19200 bps
001	

Figure 3.3 Communication settings

### 3.4 Temperature setting

In this menu, the user can change the temperature compensation type and set the temperature manually.




Figure 3.4 Temperature setting

### 3.5 Relay 1 setting

In this menu, the user can switch the relay 1 function, set the parameter alarm upper limit, alarm return difference, and alarm delay time.





Figure 3.6 Relay 1 setting

### 3.6 Relay 2 setting

In this menu, the user can switch the relay 2 function, set the parameter alarm lower limit, alarm return difference, and alarm delay time.



Function	
OFF	Low : 10.00 pH
► ON	Hys : 1.00 pH Time : 030 s

Figure 3.7 Relay 2 setting

### 3.7 Language setting

In this menu, users can change the language of the system interface. Simplified Chinese and English are built-in.

Language
► 简体中文
English

Figure 3.11 Language setting

### 3.8 Backlight setting

In this menu, the user can change the backlight mode of the LCD screen, choose whether the backlight is always on or delay off (the default is delayed off), change the backlight brightness (brightness level 1-5, brightness increase), and change the contrast.

Backlight	Mode	Parameter
► Mode	Parameter	
Mode		
► ON Delay 30s		Bright : 3 Contrast : 2

Figure 3.12 Backlight setting

### 3.9 Restore factory settings

In this menu, users can restore all current output and relay parameters to factory parameters.




Figure 3.13 Restore factory settings

#### Four, calibration

Press the "Set/Exit button" to enter the password input interface.




Figure 5 Enter the password

Enter calibration:

Enter the password "3900" to enter the calibration menu.





Figure 6 Calibration menu

#### 4.1 Automatic calibration

In this menu, users can use standard solutions to calibrate the instrument at two

points.

Auto Cal		1
6.91 pH 23 mV 25.3 °C	► 6.86	
4.08 pH 182 mV 25.3 °C	► 4.00	9.18

Figure 4.1 Automatic calibration

#### 4.2 Manual input calibration

In this menu, users can calibrate the pH of the known solution by themselves.

Manual Cal	
6.99 pH 46 mV 25.3 °C	07.00

Figure 4.2 Manual calibration

#### 4.3 Three-point calibration

Put the electrodes into the standard solution of the specified concentration in order, and press the confirm key after the data is stable.

Tri-point Cal		1
9.12 pH -172 mV 25.3 °C	► 9.18	

Tri-point Cal		2	Tri-point Cal	3
6.91 pH 23 mV 25.3 °C	► 6.86		4.08 pH 182 mV 25.3 °C	► 4.00

Figure 4.3 Three-point calibration

#### 4.4 Buffer group

**It is possible to modify the buffer standard set used in electrode calibration.**

Buffer
► JJG
NIST
4.00 6.86 9.18

Figure 4.4 Buffer group

#### **4.5 Electrode status**

You can view the current zero potential millivolt value and slope of the electrode.

Electrode
Offset : 6.7 mV
Slope : 99%

Figure 4.5 Electrode status

#### **4.5 Restore factory settings,**

The calibration parameters can be initialized to factory parameters.

Factory Reset
► Yes
No

Figure 4.5 Restore factory settings

# Appendix

## 1. Communication protocol

Communication parameters:

Baud rate: 4800, 9600, 19200 (default is 9600)

Serial data format: 8N1 (8 data bits, no parity, 1 stop bit)

Function code: 03

Device address: pH/ORP controller defaults to 1

Register definition:

Register address	Register definition	R/W	illustrate
0,1	Temp	R	°C,Float CDAB
2,3	pH	R	Float CDAB
4,5	mV	R	mV,Float CDAB
7	Temperature mode	R/W	1:ATC 2:MTC
12	RTU Address	R/W	Modbus mailing address,pH/ORP Default is 1
11	Baud rate	R/W	4800,9600,19200,The default is 9600

Detailed example of communication format:

Data read instruction:

Address + function code + register start address + register read number + CRC check code (hexadecimal)

For example Tx: 02 03 00 02 00 02 65 F8

address	function code	Register start address	Number of read registers	CRC Check code
02	03	0002	0002	65F8

Data return instruction:

Address + function code + data length + data + CRC check code (hexadecimal)

For example, Rx: 02 03 04 40 0E B8 52 4E CD

address	function	Data length	Measured value	CRC Check code

	code			
02	03	04	400EB852	4ECD

Convert the hexadecimal number 400EB852 (note the data format) to decimal by a floating-point number converter, and get the value 2.23

HEX To DEC	
HEX :	400EB852
	\$40 \$0E \$B8 \$52
DEC :	2.23000001907349

## 2. Electrode parameter table for industrial online pH controller

Electrode type	pH/ORP				ORP
Model	pH8012	pH8012F	pH8010	pH8010F	ORP8083
Range	0.00~14.00pH		-2000~2000mV		-2000~2000mV
Temp	0.0~80.0°C				
Accuracy	2%,±0.5°C		2%		2%
Pressure	0.06MPa				
IP	IP68/NEMA6P				
Theoretical slope	≥95%				

SHANGHAI BOQU INSTRUMENT COMPANY

TEL:86-021-2092-2782

WA:86-15000087545

WEB: [www.boquinstrument.com](http://www.boquinstrument.com)

Email: [michael@shboqu.com](mailto:michael@shboqu.com)

ADRESS: No. 118 Xiuyan Road,Pudong New Area,Shanghai,Zip

Code:201315,China

**\*30200134\***